

TEST CERTIFICATE

Calibrator Device Details

Serial:	2225002	Type:	TP 3M165E.2
Function:	Dry Block	Last calibrate date:	08/29/2022
Data acquisition:	Camera		
Reference:	External		

Device under test

Name:	DiTemp DT3
Serial:	540386
Type:	Pt100
Tolerance Class:	Pt100 Class B
DUT Result:	Within the tolerance

Calibration information

Ambient temperature:	21.9°C
Auditor:	Krone Demo
Calibration Date:	04/12/2023

Test Results Details

Testpoint Value[°C]	Reference Value[°C]	DUT Value[°C]	Allowed Deviation[°C]	Deviation [°C]
25.000	25.000	24.700	0.425	0.300
40.000	40.002	39.700	0.500	0.302

Testpoint 1 - Image 1



Testpoint 1 - Image 2



Testpoint 1 - Image 3



Testpoint 2 - Image 1



Testpoint 2 - Image 2



Testpoint 2 - Image 3

