



## TEST CERTIFICATE

### Calibrator Device Details

Serial:	2225002	Type:	TP 3M165E.2
Function:	Dry Block	Last calibrate date:	08/29/2022
Data acquisition:	Manual Input		
Reference:	External		

### Device under test

Name:	Demo
Serial:	123
Type:	Pt100
Tolerance Class:	Pt100 Class B
DUT Result:	Within the tolerance

### Calibration information

Ambient temperature:	22.5°C
Auditor:	Krone Demo
Calibration Date:	04/13/2023

### Test Results Details

Testpoint Value[°C]	Reference Value[°C]	DUT Value[°C]	Allowed Deviation[°C]	Deviation [°C]
25.000	24.997	25.030	0.425	0.033
50.000	50.002	50.040	0.550	0.038